

**MRI CSF Flow Foramen Magnum
with Quantification Routine Guideline (86A)**

CPT Code: 78630

Indications: Chiari malformation

Contrast: No

Sequence	Plane	Slice Thickness	Gap	FOV	Notes
T2 Blade	Sagittal	3mm	.4mm	190	
Cine_Flash 8	Sagittal	5mm	0	220	Venc 10
Cine_Flash 4	Axial	5mm	0	135	Upper
Cine_Flash 4	Axial	5mm	0	135	Lower

***POST PROCESSING ARGUS (SIEMENS)**